

TECHNICAL DATA SHEET

Extracting Forceps "79"
PROCLINIC



REFERENCES	79
MANUFACTURER	Carl Martin GmbH Neuenkamper Str. 80-86 42657 Solingen, Germany
CLASSIFICATION	Medical Devices class I, according to Directive 93/42/EWG Medical Device Regulation (EU) MDR 2017/745
DESCRIPTION	Extracting Forceps, lower jaw wisdom teeth
COMPOSITION	Surgical stainless steel
SPECIFICATIONS	Autoclavable up to 134°C. Resistant to the chemical sterilant. Supplied in "Non-sterile" condition.
PACKAGING	1 unit
INTENDED USE	Dental forceps, root remnants and splinter forceps are used to remove (extract) teeth and roots in the upper and lower jaw.
SHELF LIFE	Not applicable
SAFETY INFORMATION	Intended for use by dental professionals only. Must be checked for fault-free operation (fractures, cracks, malfunctions etc.) and be cleaned, disinfected and sterilized before each use. New instruments are contaminated with oils and fats from the manufacturing process and must therefore be cleaned thoroughly before the first sterilisation with a suitable cleaning agent.
HANDLING AND STORAGE	Sterile products must be stored in a dry, clean and dust-free environment at temperatures between 5°C and 40°C.

Solingen, 03.02.2023

Verena Gerlach

Verena Gerlach

Carl Martin GmbH
Neuenkamper Str. 80-86
D-42657 Solingen
Tel. +49 (0)212 / 81 00 44
Fax +49 (0)212 / 8 73 45
info@carlmartin.de · www.carlmartin.de

