## Huanghua Promisee Dental Co., Ltd Quality Certificate

JL-ZJ-8.2.4-03(A/0)

				JL-ZJ-8.2.4-U3(A/U)			
Product	Compule Dispenser gun		Spec.	CG01	Q'ty	150	
batch	10	Lot No.	/	Inspection basis	Product technical requirements	Inspection department	Quality inspection department
item	technical requirements		results		conclusion		
	no visible impurities and stains.			no visible impurities and stains.		OK	
appearance	The injection molding is full, consistent color, no obvious gate, no defects, cracks, bubbles, overflow edges, stains and mechanical damage.			No obvious gate, no defects, cracks, bubbles, overflow edges, stains and mechanical damage		ОК	
Connection performance	The shell case of the photocured resin container can be stuck into the slot of the working end, and cannot be stuck in place or fall off.			no failure to stick in place or fall off phenomenon		OK	
Sliding property	Hold the handle and pull the trigger to push the working end forward, and it should be smooth without clamping phenomenon.			Smooth and no jamming phenomenon		OK	
Spring return property	Hold the handle, pull and release the trigger successively, and the spring should be able to push the trigger to reset.		spring pushes the trigger back		OK		
Result	This batch of products can meet the technical requirements and can be shipped out.						
Remark							

沈慧	. 聪.	Taplati	Date:2023.5.20
Inspector:	Approved:	1	