

FICHA TÉCNICA

MEDIDOR IWANSON		
REFERENCIAS	80278	
FABRICANTE / RESPONSABLE	CHL MEDICAL SOLUTIONS, Srl Via Felice Cavallotti, 15 20122 Milán - Italia.	
CLASIFICACIÓN	Producto no Sanitario (Clase consumo).	
DESCRIPCIÓN	Instrumento para medir el espesor de las paredes de las coronas.	
ESPECIFICACIÓN	Rango de medición: 0-10 mm (medidas indicativas).	
COMPOSICIÓN	Acero inoxidable.	
VIDA ÚTIL	La vida útil del instrumento depende de la frecuencia de uso, el cuidado del usuario y los métodos de reprocesamiento adecuados.	
PRESENTACIÓN	Contiene una unidad.	
FINALIDAD	Instrumentos para coronas - para metal - calibres decimales.	
ESTERILIZACIÓN	Todos los instrumentos se deben examinar, limpiar y esterilizar antes de cada uso. Usar detergentes y desinfectantes adecuados para instrumentos médicos o dentales. El final de la vida útil de los instrumentos viene determinado por el desgaste y los daños durante el uso, y los instrumentos deben inspeccionarse para detectar defectos durante el proceso de limpieza preliminar. Esterilizar en autoclave o con calor seco.	
ALMACENAMIENTO Y TRANSPORTE	No requiere condiciones especiales de almacenamiento y transporte.	

TECHNICAL DATA SHEET

IWANSON METER		
REFERENCE	80278	
MANUFACTURER	CHL MEDICAL SOLUTIONS, Srl Via Felice Cavallotti, 15 20122 Milán - Italia.	
CLASSIFICATION	Non Sanitary Product (Consumer Class).	
DESCRIPTION	Instrument for measuring the thickness of crown walls.	
SPECIFICATION	Measuring range: 0-10 mm (indicative measurements).	
COMPOSITION	Stainless steel.	
SHELF LIFE	The lifetime of the instrument depends on frequency of use, user care and proper reprocessing methods.	
PRESENTATION	1 unit.	
FINALITY	Instruments for crowns - for metal - decimal gauges.	
STERILIZATION	All instruments should be examined, cleaned and sterilized before each use. Use detergents and disinfectants suitable for medical or dental instruments. The end of life of instruments is determined by wear and damage during use, and instruments should be inspected for defects during the preliminary cleaning process. Autoclave or dry heat sterilize.	
STORAGE AND TRANSPORT	No special storage and transport conditions required.	