

FICHA TÉCNICA

ALICATE PARA CINCHAR		
REFERENCIA	L9434	
FABRICANTE / RESPONSABLE	DENTAL WORLD S.R.L Via Antichi Pastifici,15 Zona industriale -70056 Molfetta (BA) Italia	
CLASIFICACIÓN	Producto sanitario Clase I, reglamento UE 2017/745.	
DESCRIPCIÓN	Alicate para cinchar o hacer dobleces hasta 0,7mm (.028”).	
COMPOSICIÓN	Níquel y cromo.	
PRESENTACIÓN	1 unidad por envase	
ESTERILIZACIÓN Y LIMPIEZA	Los instrumentos de ortodoncia se suministran «NO ESTÉRILES»: se recomienda lavarlos, enjuagarlos y adecuadamente esterilizados antes de cada uso. Se recomienda un aparato de ultrasonidos; si no se dispone de él, los instrumentos también pueden enjuagarse manualmente: utilice un cepillo duro con cerdas de plástico (NO cerdas metálicas) y asegúrese cerciorarse de utilizar las soluciones de lavado y descontaminación de lavado y descontaminación. Se recomienda encarecidamente que los instrumentos esterilizar el instrumental sólo después de haberlo lavado a fondo; si los instrumentos no están perfectamente limpios, la esterilización puede verse comprometida. Recuerde seguir cuidadosamente las instrucciones de uso por el fabricante del equipo de esterilización. Se recomienda la esterilización en autoclave de vapor a una temperatura aproximada entre 121°C y 134°C y una presión entre 1,03 bar y 2,3 bares.	
FINALIDAD	Su finalidad principal es proporcionar la presión necesaria para ajustar los cinchos o bandas metálicas que se colocan alrededor de los dientes de los pacientes.	
ALMACENAMIENTO Y TRANSPORTE	No requiere condiciones especiales de almacenamiento y transporte.	

TECHNICAL DATA SHEET

CINCHING PLIERS PROCLINIC		
REFERENCE	L9434	
MANUFACTURER	DENTAL WORLD S.R.L Via Antichi Pastifici,15 Zona industriale -70056 Molfetta (BA) Italia	
CLASSIFICATION	Class I medical device, EU regulation 2017/745.	
DESCRIPTION	Pliers for cinching or bending up to 0.7mm (.028").	
COMPOSITION	Nickel and chromium.	
PACKAGING	1 unit per package.	
STERILIZATION AND CLEANING	Orthodontic instruments are supplied “NONSTERILE”: it is recommended that they be washed, rinsed and properly sterilized before each use. Cleaning using an Ultrasonic device is recommended; if not available, instruments can also be rinsed manually: use a stiff brush with plastic bristles (NOT metal bristles) and be sure to use the appropriate washing and decontamination solutions. It is strongly recommended that instruments be sterilized only after they have been thoroughly washed; if the instruments are not perfectly clean, sterilization may be compromised. Please remember to carefully follow the instructions for use provided by the manufacturer of the sterilization equipment. Steam autoclave sterilization is recommended at a temperature approximately between 121°C and 134°C and a pressure between 1.03 bar and 2.3 bar.	
INTENDED USE	Its main purpose is to provide the necessary pressure to tighten the metal bands that are placed around the patient's teeth.	
INSTRUCTIONS FOR USE	See instructions on the product.	
STORAGE AND TRANSPORT	It does not require special storage and transport conditions.	