



→ IDENTIFICATION

endo revision

→ MORPHOLOGY

angled insert with 0.7 mm  $\varnothing$ , working length up to 10 mm, the last 5 mm D30 diamond coated

→ SURFACE

titanium nitride coat

→ TREATMENT

location of concealed or calcified root canal entrances, removal of restoration materials, calcification and fractured inserts in the coronal third of the root canal

