FICHA TÉCNICA

FIBRA DE REFUERZO PROCLINIC		
REFERENCIAS	7729: KIT 7741: 1MM. 7742: 2MM. 7743: 3MM. 7744: 4MM.	
FABRICANTE / RESPONSABLE	**BIOLOREN S.R.L** Via Alessandro Volta, 59 21047 Saronno - Italy	
CLASIFICACIÓN	Producto Sanitario Clase IIa, Directiva 93/42/CEE.	
DESCRIPCIÓN	La fibra de refuerzo Proclinic está compuesta de fibras blancas de polietileno de alta densidad y resistencia, totalmente compatible con composites fluidos, resinas y adhesivos. Este producto no tiene fecha de caducidad y su especial entrelazado permite un corte limpio.	
COMPOSICIÓN	Polietileno de alta densidad.	
PRESENTACIÓN	ROLLO 1 u. de 50 cm. KIT: 4 rollos de fibra x 50 cm. (1 mm. tubular, 2, 3 y 4 mm.) + tijera para cortar la fibra.	
FINALIDAD	Indicada para: Espaciadores y retenedores. Periodoncia. Fijación de puentes y prótesis. Refuerzo de dientes movidos en composite o resina. Estabilización temporal de dientes artificiales. Puentes provisionales. Retenedores de ortodoncia. Creación de muñones sobre postes.	
INSTRUCCIONES DE USO	Ver instrucciones en el producto.	
ADVERTENCIAS	No tocar Proclinic reinforcement ribbon con las manos. Proclinic reinforcement ribbon deberá estar siempre cubierto por resina de composite.	
ALMACENAMIENTO Y TRANSPORTE	No se requieren condiciones especiales de almacenamiento y transporte.	

TECHNICAL DATA SHEET

REINFORCEMENT RIBBON FIBER PROCLINIC		
REFERENCES	7729: KIT 7741: 1MM. 7742: 2MM. 7743: 3MM. 7744: 4MM.	
MANUFACTURER	**BIOLOREN S.R.L** Via Alessandro Volta, 59 21047 Saronno - Italy	
DESCRIPTION	Proclinic reinforcing fiber is composed of high-density and resistant white polyethylene fibers, fully compatible with flowable composites, resins and adhesives. This product does not have an expiration date and its special interlacing allows a clean cut.	
COMPOSITION	High density polyethylene.	
PACKAGING	ROLL 1 u. 50 cm. KIT: 4 x 50 cm fiber rolls. (1 mm. Tubular, 2, 3 and 4 mm.) + Scissors to cut the fiber	
INTENDED USE	Indicated for: Spacers and retainers. Periodontics. Fixation of bridges and prostheses. Reinforcement of teeth moved in composite or resin. Temporary stabilization of artificial teeth. Temporary bridges. Orthodontic retainers. Creation of stumps on posts.	
INSTRUCTIONS FOR USE	See instructions on the product.	
WARNINGS	Proclinic reinformcement ribbon must be covered with a resin – composite.	
HANDLING AND STORAGE CONDITIONS	No special storage and transport conditions are required.	